

Configuration summary

Mat. no.	Description Order code	Quantity	Unit
71086208	Cerabar M PMC51 PMC51-BG21JA1PGBMZJB+AB	1	ST

010	Approval:	BG	ATEX II 3G Ex ic IIC T6
020	Output:	2	4-20mA HART
030	Display, Operation:	1	LCD, push button on display/electronics
040	Housing:	J	F31 Alu, glass window
050	Electrical Connection:	A	Gland M20, IP66/68 NEMA4X/6P
070	Sensor Range:	1P	10bar/1MPa/150psi gauge, 100mH2O/333ftH2O/4000inH2O overload: 40bar/4MPa/600psi
080	Reference Accuracy:	G	Standard
090	Calibration; Unit:	B	Sensor range; mbar/bar
110	Process Connection:	MZJ	DIN11851 DN40 PN25 slotted-nut, 316L, EHEDG, 3A with seal FDA
190	Seal:	B	FKM Viton, FDA, 3A Class I, USP Class VI
500	>Additional Operation Language:	AB	German